

This item is the archived peer-reviewed author-version of:

An overview of the applied definitions and diagnostic methods to assess exercise oscillatory ventilation : a systematic review

Reference:

Cornelis Justien, Beckers Paul, Vanroy Christel, Volckaerts Tess, Vrints Christiaan, Vissers Dirk.- An overview of the applied definitions and diagnostic methods to assess exercise oscillatory ventilation : a systematic review
International journal of cardiology - ISSN 0167-5273 - (2015), p. 1-33
DOI: <http://dx.doi.org/doi:10.1016/j.ijcard.2015.04.111>

Accepted Manuscript


An overview of the applied definitions and diagnostic methods to assess exercise oscillatory ventilation – a systematic review

Justien Cornelis, Paul Beckers, Christel Vanroy, Tess Volckaerts, Christiaan Vrints, Dirk Vissers

PII: S0167-5273(15)00834-7
DOI: doi: [10.1016/j.ijcard.2015.04.111](https://doi.org/10.1016/j.ijcard.2015.04.111)
Reference: IJCA 20316

To appear in: *International Journal of Cardiology*

Received date: 10 March 2015
Accepted date: 14 April 2015



Please cite this article as: Cornelis Justien, Beckers Paul, Vanroy Christel, Volckaerts Tess, Vrints Christiaan, Vissers Dirk, An overview of the applied definitions and diagnostic methods to assess exercise oscillatory ventilation – a systematic review, *International Journal of Cardiology* (2015), doi: [10.1016/j.ijcard.2015.04.111](https://doi.org/10.1016/j.ijcard.2015.04.111)

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

An overview of the applied definitions and diagnostic methods to assess exercise oscillatory ventilation – a systematic review

Justien Cornelis^{a,1}, Paul Beckers^{a,b,c,1}, Christel Vanroy^{a,d,e,1}, Tess Volckaerts^{a,1}, Christiaan Vrints^{b,c,1}, Dirk Vissers^{a,1}

- a. University of Antwerp (Faculty of Medicine and Health Sciences), Department of Rehabilitation Sciences and Physiotherapy, Universiteitsplein 1, CDE S0.22, B-2610 Wilrijk, Belgium.
 - b. Antwerp University Hospital, Department of Cardiology, Wilrijkstraat 10, B-2650 Edegem, Belgium.
 - c. University of Antwerp (Faculty of Medicine and Health Sciences), Department of Medicine, Universiteitsplein 1, B-2610 Wilrijk, Belgium.
 - d. University of Leuven (Faculty of Kinesiology and Rehabilitation Sciences), Tervuursevest 101, B-3001 Heverlee, Belgium.
 - e. University of Antwerp (Faculty of Medicine and Health Sciences), Translational Neurosciences, Universiteitsplein 1, B-2610 Wilrijk, Belgium.
1. “This author takes responsibility for all aspects of the reliability and freedom from bias of the data presented and their discussed interpretation”

Systematic review registration number: PROSPERO CRD42013006712

Corresponding author: Justien Cornelis

University of Antwerp (Faculty of Medicine and Health Sciences), Department of Rehabilitation Sciences and Physiotherapy, Universiteitsplein 1, CDE S0.22, B-2610 Wilrijk, Belgium. Email: justien.cornelis@uantwerpen.be, Tel: +3236418265, Fax: +3236418270

Acknowledgment of grant support: None

Conflicts of interest: None

Keywords: Cardiopulmonary exercise, Humans, Oscillatory ventilation, Definition, Assessment, Review

Abstract

The variable “exercise oscillatory ventilation” (EOV), assessed during cardiopulmonary exercise test (CPET), recently became a fundamental prognostic parameter in patients with heart failure. In literature, various definitions are suggested, but an uniformly accepted description to identify EOV still lacks. We performed a systematic review of the literature in order to determine the different definitions and diagnostic techniques to assess EOV. A systematic search strategy was established and executed in seven databases (PubMed, Google Scholar, Cochrane Clinical Trials, Science Direct, Pedro, Web Of Science library and Medline (Ovid)) resulting in 605 citations after de-duplication. Full-text articles (n=124) were assessed for eligibility, resulting in 75 citations. The review accounted 17440 patients of whom 4638 subjects presented EOV. Seven studies described EOV in a non-heart failure population accounting 168 EOV subjects. The definitions could be categorized in nine subdivisions of which four (n=43) referred to an original description. The other subdivisions were combinations of the original definitions (n=11), quantifications (n=4), computational (n=3), vaguely described (n=8) or not defined (n=6). Symptom limited maximal exercise tests were conducted to assess EOV, however the modes, protocols, software and data sampling were divers. Heterogeneity in the numerous definitions to identify EOV and the vaguely described assessment methods are hindering the evolution to a standardized uniformly accepted definition and technique to identify this abnormal breathing pattern. Unity in definition and international adopted assessment is warranted to strengthen its validity as a prognostic marker and could promote communication. It may facilitate clinical trials on pathophysiology and origin of EOV.

Keywords: Cardiopulmonary exercise, Humans, Oscillatory ventilation, Definition, Assessment, Review

1. Introduction

In a population with heart failure (HF), exercise oscillatory ventilation (EOV) emerges as a key prognostic parameter that tends to outperform the formerly suggested predictive cardiopulmonary exercise test (CPET) derived variables such as peak oxygen uptake (VO_2) and ventilation over carbon dioxide (VE/VCO_2) slope[1,2]. Furthermore, this oscillatory breathing pattern indicates general deterioration of other CPET derived prognostic values[2]. The reported incidence of this waxing and waning pattern in ventilation (VE) during exercise is 7 to 51%[3,4]. This broad interval originates as no universal definition is currently described to assess EOV[5,6]. The lack of a gold standard definition could arise from the wide variability within the EOV pattern characteristics e.g. amplitude, duration of the oscillations and period[7,8]. Moreover, assessment of EOV is mainly done by manual scoring or visual interpretation of at least one clinician, which is subjective and could therefore vary considerably between centres and individuals[8]. In 2010, the American Heart Association (AHA)[1] declared the definition adopted by most studies stated that oscillations were present for at least 60% of the exercise time with an amplitude $\geq 15\%$ of the average resting value, in accordance to Corrà et al.[9]. This description was followed by a joint scientific statement of the AHA and the European Society of Cardiology (ESC)[5]. Moreover, many different CPET systems are available with each its unique software for processing, analysing and displaying data[1]. This software also determines data sampling and averaging, which the user can select. Modern CPET systems provide a wide variety of automated reports, causing a selection of the available data for the clinicians' interpretation and therefore significantly influencing the assessment of the test results and the data analysis. It is stated that there is an important need of standardization in data sampling e.g. breath by breath (BbB), the amount of breaths, moving averages or specific time intervals, as there is a misbalance in these methodologies between precision and variability. Especially in EOV, of which the assessment is based on pattern recognition, relates the difficulty in visualization of the signal for analysis. Visualization is commonly done by using a moving average filter which smooth the data and eliminate small variations[10]. The ESC and AHA council recommends to use 10s averaged ventilation data for plotting[5]. Furthermore, each stadia of the test e.g. rest, warm-up, exercise and cool-down should be reported as well as the selected exercise protocol as both could influence the interpreted variables[1]. In a recent meta-analysis of our research group[2], it was stated that more research should be done on defining and assessing EOV in a more accurate and reproducible way. We found that difference in defining EOV influenced significantly the results. Despite, the mode of protocol was an independent parameter, all sub-group analyses were characterized by a high overall

heterogeneity. Therefore we performed a systematic search strategy to provide an overview of the suggested definitions and assessment techniques used to analyse EOV. The purpose of this systematic review was to identify all the applied definitions of EOV through a symptom-limited exercise test (SLMT) and to summate the described test characteristics in literature independently from pathology. As the pathophysiology of EOV is not yet fully understood [11,7], analysis of EOV in other populations than HF could lead to new insights in this aspect.

2. Methods

A systematic search strategy was established: “exercise oscillatory ventilation” OR “exercise periodic breathing” OR “exertional periodic breathing” OR “exertional oscillatory ventilation” OR “exercise oscillatory breathing” OR “exertional oscillatory breathing”. In the Pedro database only “exercise oscillatory ventilation” was searched as the other quotations were not found. As indicated in the specific database, the following filters were applied: only humans (PubMed and Science Direct) and journals (Science Direct), no reviews (Cochrane), no patents or citations (Google Scholar).

The following inclusion criteria were a priori established: (1) defining and/or investigating oscillatory minute VE during a CPET; (2) all populations; (3) all interventions; (4) humans; (5) English, French, German, Dutch, Spanish or Portuguese; (6) normal environmental conditions. The exclusion criteria were: (1) oscillations described during rest, sleep and/or steady-state exercise; (2) the study designs were reviews, meta-analyses or books; (3) high altitude. Expert opinions and recommendations on on-going studies or other relevant data were gathered. Conference proceedings, index lists, doctoral manuscripts, commentaries and editorials were screened and included if oscillations in minute VE were defined through exercise and if it concerned original research by the authors. Reference lists were checked for any topic-related studies. The corresponding author of a study was contacted if needed to obtain any missing information or data. If authors could not be reached or if the data was not available, the citation was excluded. The search included citations published up until October 24th 2014.

All citations identified by the above mentioned electronic databases were organized and the duplicates were deleted. Initially, two investigators (J.C. & T.V.) screened the results from the electronic searches in order to select potentially relevant citations based on titles and abstracts. Full-text articles were retrieved and evaluated based on the proposed in- and exclusion criteria. In case of uncertainty, a third investigator (C.V.) evaluated the citation and

consensus was sought during a meeting. Although the authors were responsible for a specific step in the screening process, the final determination on whether an article was included or excluded was based on common agreement.

The following study characteristics were extracted, and if appropriate coded, from the articles by two researchers (J.C. & C.V.). Concerning the assessment of EOv by CPET, the variables rest, warm-up, mode, protocol, intensity, data sampling and data collection were retrieved. Furthermore, the described EOv definition was acquired. Secondary collected data concerned “population, study centre, respiratory exchange ratio (RER), pathophysiology and incidence of EOv”.

Methodological scoring nor Risk of Bias were performed as we only extracted data concerning the methodological analysis and applied definitions in order to evaluate EOv. This scoring would not give an additional value regarding the studies included in this review.

3. Results

3.1 Selection of the studies

Seven databases (PubMed, Google Scholar, Cochrane Clinical Trials, Science Direct, Pedro, Web Of Science library and Medline (Ovid)) were consecutively systematically searched, resulting in 751 citations. Five studies[12,13,3,14,15] were identified after manual reference list screening, resulting in 756 citations. After de-duplication, 605 articles were screened for eligibility on title and abstract, resulting in 124 citations. Ultimately, full text screening resulted in 75 studies (Figure 1).

[Insert Figure 1]

Since the group of Sun et al. reported on the same sample in two different studies, description of the population was only mentioned once[4,16].

3.2 Characteristics of the included studies

The 75 studies included in this review, accounted for 17440 patients of which 4638 (26.6%) presented EOv (Figure 2). Seven of these included studies described EOv in another population than HF e.g. cardiac dysfunction[17,18,14], idiopathic dilated cardiomyopathy[19], total cavopulmonary connection after Fontan[20], liver transplantation[21]

and congenital cardiac disorder[10], accounting 752 subjects of which 168 presented EOv. EOv was not reported in healthy control patients[12,22-27]. Ten of the included studies were conducted in a multicentre setting [4,16,28-35], totalling 113 centres, representing 8326 subjects with HF of which 2150 presented EOv. The lowest and highest reported incidence rate was 7% [36,37] and 64% [38], respectively.

[Insert Figure 2]

3.3 Assessment

Assessment through a SLMT was performed in all included studies, yet characteristics were only described in 68 citations (n=68) (Table 1). The RER varied between 1.00[33,39] and 1.10[40,21], mainly set ≥ 1.05 [35,9,29] for inclusion. The selected mode of protocol in these studies was bicycle (n=38), treadmill (n=11) or a combination of both modes (n=7). The passive rest, active (un)loaded warm-up and the recovery phase during one to three minutes after SLMT was registered in 18, 25 and five articles, respectively[41-44,19]. The (modified) ramp protocol was selected in the majority of the studies using a bicycle. In treadmill, a (modified) Bruce protocol was often selected. Twenty-two different CPET devices of 12 manufactures were used to collect the data. Generally, BbB data were collected. In order to assess EOv, a moving or fixed averaging method was often applied. In the majority of the articles, SLMT was conducted during 8-12min. In a few articles, other limits were obtained such as 7-10min[45] and 6-15min[4,16].

[Insert Table 1]

3.4 Definitions

The definitions described in the included studies (n=75) could be categorized in nine subdivisions of which four (n=44) referred to an original definition as suggested by Kremser et al.[12] (n=23), Leite et al.[23] (n=13), Ben-Dov et al.[46] (n=6) and Sun et al.[4] (n=2) (Table 2; Figure 2). The other subdivisions were combinations of the original definitions (n=11), quantifications (n=4), computational (n=3), vaguely defined (n=8) or undefined (n=6). One study[47] compared the difference in EOv incidence upon two definitions[46,9]. When the reported original definitions were applied, EOv was visually determined[21,48,37,49,50] or quite rapidly manual calculated [32,51,52] by at least one clinician. In seven studies, EOv was determined by quantification or by computational techniques. One study [53] reported the use of a familiarization test to confirm the presence of EOv.

[Insert Table 2]

4. Discussion

To our knowledge, this systematic review is the first to provide a full overview of the applied definitions and diagnostic methods to assess EOv. An important finding is the wide variety of definitions applied to describe EOv. Four original descriptions with many modifications were suggested[12,23,46,4]. Attempts to quantify EOv[24,13,15,54] without proposed cut-off value were performed and authors aimed to assess this abnormal breathing pattern with specific computer techniques[55,25,10]. Moreover, numerous CPET protocols, data collection approaches and data sampling procedures were applied in the included studies.

Exercise oscillatory ventilation is especially recognized in HF, however this review indicates EOv can appear in other cardiac related populations[17,18,14,19,20,10]. The most striking finding is the presence of EOv in liver transplant patients[21]. However, it is not clear whether the subjects presenting EOv in that study had no underlying cardiac pathology. It could be assumed that EOv is a patient related pathophysiologic indicator as it was not mentioned in normal subjects yet[22,26].

To our knowledge, the first authors to describe the oscillations in ventilation were Kremser et al. in 1987[12]. Since then, investigators tried to discover the best possible way to define and assess those oscillations. Up until now, no gold standard was proposed[11], causing a wide variety of suggested definitions for EOv. Recently, the AHA and ESC council[1,5] published guidelines in which the definition by Corrà et al.[9] was elected to define EOv. In table 2 is shown that this definition is a modified description of the proposed criteria by Kremser et al.[12]. Only one third of the citations included in this review, used this description to assess EOv. It is unclear why the guidelines suggest that definition in particular. Furthermore, it remains elusive why the majority of the authors refer to Corrà et al.[9] and only a few to Kremser et al.[12], although the latter was published 15 years earlier. Furthermore, it seems that two definitions were used in Corrà et al.[9,48]. After contacting the authors[9], they confirmed that the second description (>60% duration; amplitude >15%) is the one to follow, not corresponding to the suggested definition in the AHA and ESC guidelines ($\geq 60\%$ duration; amplitude $\geq 15\%$). We could assume that the proposed definition in the guidelines slightly misquoted the original author.

Lately, Sun et al.[4] analysed oscillations in other parameters in addition to VE in order to define EOv. However, it was stated[25] that the oscillations in $\dot{V}O_2$, $\dot{V}CO_2$ and RER were proportionally smaller than those in VE, and were also different in phase. Therefore, it is questionable if oscillations in other cardiorespiratory parameters simulate the same underlying pathophysiologic processes.

Apart from the wide variety of definitions also many different interpretations were mentioned concerning the mathematical terms and characteristics within the described definitions. Regarding the cycle length, many descriptions were given such as the distance between two nadirs[26], the period of observed oscillation divided by the number of oscillation[56] and the interval from peak to peak for each cycle at rest expressed as mean value[18]. Moreover, different values were proposed for the duration of the cycle length such as approximately 60s [57], 40s-140s[4,16], 30s-60s[23] or a percentage of the average[46]. Furthermore, the total duration of the oscillations could also vary. Suggested durations were $>66\%$ of the exercise protocol [12], $>$ or $\geq 60\%$ [9], $\geq 50\%$ [58], 2 or 3 regular oscillations[46,23]. Sometimes these characterizations were measured at a specific point in time such as at 50% peak exercise[38,36]. Concerning the amplitude, numerous descriptions were suggested. The amplitude was described as the difference between the peak VE of the oscillation and the average of the VE of the two surrounding nadirs[26]; by calculating the variation coefficient of VE[35]; by assessing the correlation coefficient of VE[36]; or expressed as a percentage of the mean of the rest VE[18]. In addition, “a nadir”, mentioned in numerous definitions[26,17], is neither a mathematical nor a medical term, but rather a term used in physics. We assumed to interpret this term as local minima. A recent meta-analysis[2] stated that presence of EOv depended upon the proposed definition. Therefore it is important to describe these above mentioned sub-characteristics unambiguously.

Regarding computational techniques to define EOv, there is an uncertainty about the interpretation of the signal using a Fourier transformation[8,25] as the researchers assumed EOv is a stationary signal. However, other investigators[10] indicated that EOv is a non-stationary signal and therefore Fourier would not be the ideal transformation to apply in a clinical individualized setting as assessment in a frequency domain is not useful for clinical application. These authors applied a calculation of the mean BbB variability in respiratory gas exchange[10]. These developed software algorithms are promising to bring new insights in the analysis of EOv. We assume that up until now no clear detection methods other than visualization during data collection or chart review were effectively applied. These techniques are characterized by a high heterogeneity as medical charts typically only provide

summarized information and hand scoring could vary considerably between individuals and centres. Moreover, some authors tried to quantify the oscillations and thus obtained to describe an amount, recognized by a number[13,24,54]. However, no clear cut-off values to quantify the oscillations were suggested. Yet, the authors [54] make clear that it is a different approach compared to the other studies in which the presence of EOV is analysed by YES/NO criteria. In general, more research towards defining EOV is mandatory.

In the majority of the included studies, data sampling techniques were described. Twelve different unique software systems, of which a few are validated[14], were applied to register and analyse the data. In literature is stated that those devices are not always interchangeably in order to assess peak VO_2 and VE [59]. This attest the possibility that towards analysing EOV, there is a lack of congruence between the devices, contributing to a higher heterogeneity[1]. Furthermore, it was suggested that both in clinical practice and in research publications data from patients ought to be accompanied by details of the averaging technique used[1]. This was stated as important as knowing whether the exercise is on a treadmill or a bicycle and the protocol used. This information would assist in interpreting discrepancies between findings from different centres[25]. In the majority of the studies presented, the data was recorded BbB. Consequently a data averaging method was often reported. These researchers averaged the data assumingly out of practical consideration or to visualize the oscillations more. These features indicate a high heterogeneity in data sampling techniques which could lead to a higher variability and lower precision in assessing EOV[1]. The recommended[5] 10s averaging VE data technique was only applied in ten of the included studies. However, it is unclear why this would be the advised data sampling technique in order to estimate EOV. The effects of data averaging (using different periods from 10s to 90s) on the size of the oscillations as well as on the noise-to-signal ratio was already investigated[25], recommending a standardized averaging period similar to the length of the typical periodic breathing cycle, i.e. 60s regardless of whether breathing is periodic or stable. These authors suggested this would improve the physiological validity of the value obtained in subjects with periodic breathing, without adverse effects in subjects with stable breathing. Moreover, they assumed that ideally assessment of EOV should be made on raw data before 60s averaging, as filters eliminate small variations in gas exchange parameters and clarifies the trend in a set of data [25,10]. Since in practice the exact period of oscillatory breathing may vary between patients, no single rectangular window can be guaranteed to eliminate oscillations in metabolic parameters[25]. Furthermore, it is important to evaluate this phenomenon objectively and therefore preferable without excluding these small variations in gas exchange[10].

In the included citations, often large databases were screened over many years. A variety on CPET protocols and data collection methods were suggested. The adopted modes in the SLMT were generally reported. Bicycle as well as treadmill were applied. In the multicentre studies included in this review, those two modes were often used interchangeably. Some investigators declared the results were correctly registered irrespective the mode of exercise[30,31]. However, the EOV prevalence was 25% among patients tested on a bike and 41.3% among those tested on a treadmill[31]. In future studies, both modes could be compared in the same sample of patients to accurately investigate the difference in EOV incidence. Concerning the applied protocol, it was expressed that each stadia of the exercise test (rest, warm-up, exercise and recovery) should be reported in a study assessing cardiac patients[1]. Especially in EOV, this data is of great importance to the authors [12,9], as the investigators defined EOV by comparing oscillations in rest data before the actual exercise test. However, this data was rarely mentioned in the included studies. One can argue if oscillations during rest should be classified as periodic breathing or as exercise oscillatory ventilation. Therefore, it is unclear if defining EOV requires analysis of the oscillations during rest or only during warm-up and exercise[60]. However, it could be interesting to evaluate the oscillations during recovery, as assessed in other prognostic parameters after submaximal exercise testing[61]. Many researchers stated and included the clinical test as maximal if a RER of 1.05 or even 1.00 was reached. However the guidelines described $RER \geq 1.10$ as cut-off[1]. One could state that EOV can be assessed by submaximal exercise testing, making the RER value inferior to other characteristics of the exercise test. However, even this should be standardized as Scardovi et al[39] presented a high percentage of EOV in female patients with HF using a 1.00 cut-off for RER, while in the majority of the included articles more male subjects were described. These features confirmed that despite the unpredictability of EOV, there is a need of standardized CPET protocols and data collection to assess EOV in prospective conducted studies.

4.1 Study limitations

We obtained to categorize the wide variety on descriptions, mentioned in literature to define EOV, in nine sub groups to provide a clear overview. In order to create these sub groups, the definitions were classified as modified or originally stated. Whether the definition was a modified or an original description was based on the publication date, the references made to the applied definition in the articles and the personal interpretation by two researchers (J.C.

and C.V). However, the nine sub groups in definition provide a clear general overview of all descriptions stated in literature, a general definition for EOv is mandatory. This overview does not provide a gold standard definition.

The lack of transparency towards assessment, the major variation on data sampling techniques and the diversity of the conducted studies, does not allow us to make a clear statement on which mode, protocol and data sampling technique should be used to assess a patient with EOv. Further research towards defining and assessing EOv is required.

5. Conclusions

Heterogeneity in the numerous definitions to identify EOv and the vaguely described assessment methods are hindering the evolution to a standardized uniformly accepted definition and technique to identify this abnormal breathing pattern. Unity in definition and international adopted assessment is warranted to strengthen its validity as a prognostic marker and could promote communication. It may facilitate clinical trials on pathophysiology and origin of EOv.

Acknowledgment of grant support: None

Conflicts of interest: None

Table 1 An overview of the applied symptom limited maximal test (SLMT) characteristics and data sampling.

Author (year)	Mode	Rest (min)	Warm-up		Protocol; Intensity	Data collection (Averaging); Software
			Time (min)	Load (W); rpm		
Agostoni[22]	B	3	3	0	Ramp	BbB; Vmax ^a
Agostoni[53]	B	-	1	0; 60	Continue ↑6-12W/min 60rpm	BbB; Vmax ^a
Agostoni[28]	BT	-	-	-	Ramp (B) & Modified Bruce (T)	BbB
Agostoni[29]	BT	-	-	-	Ramp	BbB
Apostolo[62]	B	>5	>3	0	Ramp; ↑5-10W/min	BbB; Vmax ^a
Arena[63]	T	-	-	-	Conservative ramp; ↑±2mlO ₂ /kg/min in workload every 30s	BbB (10s)
Arena[64]	T	Yes	-	-	Conservative treadmill protocol	BbB (10s); Vmax29 ^a
Ben-Dov[23]	BT	Short	2-3	0 (B)	Naughton modified ↑ every 2min from 1mph 0% to maximum 3.5mph 16% or Ramp ↑10-15W/min or ↑ in 1min steps	BbB (8 breath moving); System 2001 ^b or CPX ^b
Cahalin[41,65]	B	-	-	-	Ramp	BbB (30s, printed rolling averages 10s); Vmax ^a
Callegaro[54]	T	-	-	-	Ramp; 2.4km/h 2% with ↑ every 20s in speed and slope	BbB; Metalyzer3B ^c
Castro[66]	B	-	-	-	↑5W every minute 60rpm	BbB; Vmax229 ^a
Corrà[9,48,37,49]	B	-	1	0	Ramp; 10W/min 60rpm	BbB; Vmax29 ^a
Corrà[67]	B	2-3	1	0	Ramp; 10W/min 60rpm	BbB; Vmax29 ^a
Dall'Ago[13]	T	-	-	-	Ramp; 2.4km/h 2%, ↑ every 20s of speed 0.1-0.2km/h and ↑ every 60s 0.5%-1.0% in slope	(20s mean aliquots); TEEM 100 ^d
Feld[3]	-	-	-	-	-	(8-breaths confirmed by converting data to BbB); System 2001 ^b
Francis[24]	T	Short	-	-	Bruce; Modified initial stage 0.5mph 5%	BbB on-line; Amis 2000 ^e
Francis[25]	T	-	-	-	Bruce; Modified initial stage 1mph 5%	BbB on-line (different moving windows 15s, 30s, 60s); Amis 2000 ^e
Guazzi[30]	BT	-	-	-	Ramp	BbB (30s, printed in rolling averages 10s)
Guazzi[68-70]	B	-	-	-	Ramp	VE collected continuously; CPX-D ^b

Guazzi[31]	BT	-	-	-	Ramp	BbB (30s, printed in rolling averages 10s); CPX-D ^b or Vmax ^a
Guazzi[71]	B	-	-	-	Ramp	BbB (30s, printed in rolling averages 10s); Vmax ^a
Guazzi[32]	BT	-	-	-	Ramp	BbB (30s, printed in rolling averages 10s); CPX-D ^b & Vmax29 ^a
Guazzi[72]	B	-	-	-	Ramp	VE collected continuously at rest and throughout exercise; CPX-D ^b
Guazzi[51,52]	B	-	-	-	Ramp	BbB (30s, printed in rolling averages 10s); Vmax29 ^a
Guazzi[43]	B	3	-	-	Ramp	BbB (10s); Vmax Spectra ^a
Guazzi[73]	B	-	2	0; 60	Ramp	BbB (30s, printed in rolling averages 10s); Vmax29 ^a
Ingle[47,60]	T	-	-	-	Bruce; Modified initial stage 2.74km/h 0%	Oxycon Delta ^f
Joho[56]	B	-	3	0	Progressive ↑5-10W/min	BbB (EOV cycles displayed on an expanded time scale); AE-300 ^g
Kato[18]	B	4	4	0 or 20; 60	↑1W/6s (10W/min)	BbB (interpolated to SbS values; calculated as successive 3s averages translated to 5 point moving average); AE-300s ^g
Kazimierczak[45]	B	-	-	-	Ramp	BbB (10s); ZAN680 ^h
Koike[17]	B	4	4	10 or 20	↑1W/6s (10W/min) 60rpm	BbB 4min before start and until end of exercise (5 point moving); AE-280 ^g
Kremser[12]	B	-	3-4	0	Continuous ↑6-10W/min or at 1min interval 20-100rpm	BbB (30s; breaths with a tidal volume <250ml were not analyzed)
Lasso[44]	B	3	3	0	Naughton & San Ignacio 20 ↑20W/min	Vacu Med-Vista mini CPX (software Turbofit 5.12) ^b
Leite[46]	B	-	-	-	Ramp	BbB (8 breaths moving average); CPX ^b
Lim[74]	T	-	-	-	Ramp	Powercube ⁱ
Matsuki[75]	B	yes	3	-	Ramp; 10W/min	AE310 ^g
Miyagi[14]	B	-	3	0	↑3-15W/min	BbB; AE280 ^g
Muneuchi[20]	T	-	3	0	↑15%/min	BbB
Murphy[26]	B	-	3	0	Ramp; continuous ↑5-15W/min	BbB (1min); Medgraphics ^b

Neviere[21]	B	2	2	0	Ramp; 10W/min	BbB (10s moving); Medisoft ErgoCard ^j
Okumura[19]	B	3	3	<10	Ramp; 10W/min	-
Oldenburg[76]	B	-	-	-	10W and continuous ↑10W/min	BbB (1min)
Olson[36]	T	-	-	-	2.0mph 0%; ↑ every 2min to yield an approximate 2MET per work level	BbB (averages obtained over the final 30s of each workload); Medgraphics ^b
Olson[8]	B	10	-	-	-	BbB; CPX/D ^b
Ribeiro[77]	B	-	-	-	0W ↑15W every 3min 60rpm	BbB; Sensormedics Inc ^a
Roche[40]	B	2	-	-	10W ↑10W progressively 60rpm	(Every 15s); ExpAir Soft ^j
Sato[78,79]	B	-	-	-	Ramp	BbB; AE-300S Aeromonitor ^e
Scardovi[33,39]	B	-	-	-	Ramp	BbB (10s intervals); Vmax29 ^a
Schmid[34]	B	-	3	0; 60	Ramp or ramp equivalent (load ↑/min; 0W ↑6-12W/min)	BbB (10s); Vmax ^a or Oxycon Alpha ^k
Sperandio[42]	B	-	3	0; 50±5	Ramp; 5-10W/min 50±5rpm	BbB (arithmetic mean of 20s); CardiO ₂ System ^b
Sun[4,16]	BT	3	3	0	Ramp; 5-15W/min (usually 10W/min) or 1min step intervals	BbB (10s bins); 6 different systems
Ueshima[50]	B	3	3	10	Ramp; 15W/min	AE-280 ^e
Vangesselen[10]	T	-	-	-	4.8km/h <6years or 5,6km/h >6years; 0% ↑2%/min; severe cardiac impairment adapted protocol initial 1km/h/min ↑2%/min	-
Wang[57]	B	-	2	0	Continuous ↑10W/min	BbB; MasterScreen CPX ^l
Winkelmann[15]	B	-	-	-	-	BbB; Metalyzer3B ^c
Yajima[27]	B	4	3	20;60	↑1W/6s	BbB (5 breath moving); RM-300 ^e
Zurek[35]	B	yes	3	0	Ramp	BbB (rest & warm-up 60s; exercise 30s); Oxycon Alpha ^k & Vmax29C ^a

Bicycle (B); Treadmill (T); Bicycle & Treadmill (BT); minute (min); watt (W); rates per minute (rpm); second (s); Breath-by-Breath (BbB); ventilation (VE); Second-by-Second (SbS). ^a SensorMedics, Yorba Linda, CA, USA or SensorMedics, Homestead, FL, USA; ^b Medical Graphics Corporation, St Paul, MN, USA or Medgraphics, Minneapolis, MN, USA; ^c Cortex, Leipzig, Germany; ^d Aerosport, Ann Arbor, Michigan, USA; ^e Innovision, Odense, Denmark; ^f ViaSys Healthcare Inc, Philadelphia, Pennsylvania, USA; ^g Minato Medical Science, Osaka, Japan; ^h ZAN Mengerate, GmbH, Germany; ⁱ Ganshorn, Medizin Electronic, Neuenkirchen, Germany; ^j Medisoft, Dinant, Belgium; ^k Jaeger-Toennies, Höchberg, Germany; ^l Cardinal-Health, Germany.

Table 2: An overview of the different descriptions used to define exercise oscillatory ventilation (EOV).

Description	Citations quoting the description
Original definition of Kremser et al. (1987)[12]: Ventilatory oscillations lasting >66% of the exercise protocol, with an amplitude >15% of the average value at rest.	
Modified >60% duration; amplitude >15%	[9,48,37,49,47,19,20,80]
$\geq 60\%$ duration; amplitude $\geq 15\%$	[9,48,74,63,64,52,73,32,69,70,51,72,81,75,82]
Cyclic fluctuations in minute VE at rest that persist during effort $\geq 50\%$ of the exercise duration	[58]
Original definition of Leite et al. (2003)[46]: ≥ 3 regular oscillations (i.e. clearly discernible from inherent data noise) in VE; Regularity was defined if the SD of 3 consecutive cycle lengths (time between 2 consecutive nadirs) was within 20% of the average; minimal average amplitude of ventilatory oscillation $\geq 5l$ (peak value minus the average of 2 in-between consecutive nadirs).	[47,33,35,21,30,68,45,60,57,56]
Modified ≥ 3 regular oscillations in ventilation	[83]
Amplitude of oscillation >30% of the mean value for VE, VO ₂ or VCO ₂ , ≥ 3 consecutive oscillations with amplitude of $\geq 5L/min$ for VE and $\geq 3ml/kg/min$ for VO ₂ or VCO ₂ .	[40]
Original definition of Ben-Dov et al. (1992)[23]: Marked oscillations of 30-60s duration. Magnitude (Δ) VE ($\Delta = (\text{peak-nadir})/\text{mean}$ over the time period of the oscillation) $\geq 25\%$ in ≥ 2 consecutive cycles (nadir to nadir) during exercise visually and normalized to 360°.	
Modified Regular oscillations of VE amplitude >25% of mean amplitude VE.	[38]
≥ 2 consecutive cycles of clear ventilatory oscillations at rest or warm-up period. The mean of the differences between the peak and nadir of oscillating VE is >30% of the mean value of VE.	[50,17]
≥ 3 consecutive cycles of clear ventilatory oscillations were noted from the beginning of rest until the end of the warm up. Thereafter, the mean of the differences (amplitudes) between the peak and nadir of oscillating VE at rest and the mean resting VE was calculated.	[18]
Amplitude was >40% of mean VE.	
Clear regular cyclic waxing and waning of VE without apnea, an amplitude of VE >30% of the mean VE at cycle peak and ≥ 3 consecutive cycles at rest and during exercise. The same 3 cycles were used for the measurement of amplitude (the difference between	[8]

peak and nadir of a cycle and averaged over 3 consecutive cycles) and period (the time interval between 2 consecutive nadirs or 2 consecutive peaks and averaged over 3 consecutive cycles). Quantifications were also obtained and Fourier analysis was also conducted.

Original definition of Sun et al. (2010)[4]: ≥ 3 consecutive cyclic fluctuations of VE. To be defined as positive, the amplitude of oscillatory ventilation $>30\%$ of concurrent mean VE with a complete oscillatory cycle within 40s-140s. Oscillations of similar frequency must also be visible in ≥ 3 or more of the following variables: oxygen pulse, VO_2 , VCO_2 , VE/VCO_2 , RER (i.e. VCO_{2e}/VO_{2e}), $PetO_2$ or $PetCO_2$.

Modified ≥ 3 consecutive cyclic fluctuations of VE or $\geq 60\%$ of entire CPET period. [16]

Combinations of the above stated definitions

≥ 3 regular oscillatory fluctuations in VE, with minimal average amplitude of $\geq 5l/min$ persisting for $\geq 60\%$ of the entire exercise time. [65,41,71]

$>60\%$ duration; amplitude of $>30\%$ of the mean VE. [34,53]

Oscillations $\geq 60\%$ of entire exercise data at an amplitude $>15\%$; amplitude of VE $\geq 5l/min$; a regular oscillation as defined by a SD of 3 consecutive cycle lengths within 20% of the average. [31,43,42]

Cyclic fluctuation of VE at rest and during exercise with amplitude swings $>30\%$ of the mean VE, $>15\%$ for $\geq 60\%$ of incremental exercise duration. [62]

≥ 3 consecutive, regular oscillations in VE during exercise with oscillation amplitude $\geq 25\%$ of average VE, persisting for $\geq 60\%$ of exercise duration. [26]

$\geq 25\%$ variation of the amplitude of VE, persisting for $\geq 60\%$ of exercise duration. [36]

Quantifying oscillations

Oscillatory breathing at rest and during CPET. For each cycle of oscillation, the mean and amplitude (half the difference between peak and trough) [24]

was calculated. The relative amplitude (α) as the ratio between amplitude and mean e.g. if VE varied between 15 and 25l/min, the mean 20l/min was defined; than amplitude would be 5l/min and αVE would be $5/20 = 0.25$

The relative amplitudes of oscillations (α) in VE, VO_2 and VCO_2 were calculated for every 20-s period as the ratio between amplitude and its respective mean throughout the test. [13]

For every 2 adjacent 20-s period of VE, the amplitude of oscillation was calculated as difference between the 2 points divided by their mean. This value was again divided by the mean to obtain the relative amplitude, and the values of the entire test were averaged to convey in a single ratio. [15,54]

Computational

- Computational analysis using the theorem of Fourier. [55,25]
- Computational analysis calculating the mean BbB variability in respiratory variables. [10]

Vaguely defined (referred to)

- Exercise-induced periodic oscillation in VE, VO₂, VCO₂ and the RER. [22]
- Cyclic fluctuations of ventilation.([48]) [28,29]
- Oscillatory ventilation during exercise is characterized by a TV that acts as a crescendo-decrescendo pattern without apnoea period.([68,46,4,84,85]) [44]
- Visual oscillations. [77]
- Ventilatory oscillations.([23]) [27]
- Cyclic pattern of respiration confirmed on two separate occasions in each patient. [3]
- An oscillatory hyperventilaton pattern. [14]

Not defined (referred to)

Scardovi[39] ([46]); Oldenburg[76] ([9]); Castro[66] ([46,47]); Corrà[67] ([12]); Sato[78,79]

minute ventilation (VE); standard deviation (SD); oxygen uptake (VO₂); carbon dioxide output (VCO₂); respiratory exchanger ratio (RER); ventilation over carbon dioxide (VE/VCO₂); litre per minute (l/min); second (s); Breath-by-Breath (BbB); tidal volume (TV); percentage (%); pulmonary end-tidal oxygen uptake (PetO₂); pulmonary end-tidal carbon dioxide output (PetCO₂); cardiopulmonary exercise test (CPET)

References

1. Balady GJ, Arena R, Sietsema K, Myers J, Coke L, Fletcher GF, Forman D, Franklin B, Guazzi M, Gulati M, Keteyian SJ, Lavie CJ, Macko R, Mancini D, Milani RV, Amer Heart Assoc Exercise C, Council Epidemiology P, Council Peripheral Vasc D, Interdisciplinary Council Q (2010) Clinician's Guide to Cardiopulmonary Exercise Testing in Adults A Scientific Statement From the American Heart Association. *Circulation* 122 (2):191-225. doi:10.1161/CIR.0b013e3181e52e69
2. Cornelis J, Taeymans J, Hens W, Beckers P, Vrints C, Vissers D (2015) Prognostic respiratory parameters in heart failure patients with and without exercise oscillatory ventilation — A systematic review and descriptive meta-analysis. *International journal of cardiology* 182 (0):476-486. doi:http://dx.doi.org/10.1016/j.ijcard.2015.01.029
3. Feld H, Priest S (1993) A cyclic breathing pattern in patients with poor left ventricular function and compensated heart failure: A mild form of cheyne-stokes respiration? *Journal of the American College of Cardiology* 21 (4):971-974. doi:http://dx.doi.org/10.1016/0735-1097(93)90355-5
4. Sun XG, Hansen JE, Beshai JF, Wasserman K (2010) Oscillatory breathing and exercise gas exchange abnormalities prognosticate early mortality and morbidity in heart failure. *Journal of the American College of Cardiology* 55 (17):1814-1823. doi:10.1016/j.jacc.2009.10.075
5. Guazzi M, Adams V, Conraads V, Halle M, Mezzani A, Vanhees L, Arena R, Fletcher GF, Forman DE, Kitzman DW, others (2012) Clinical recommendations for cardiopulmonary exercise testing data assessment in specific patient populations. *European heart journal* 33:2917-2927
6. Pardaens S, Calders P, Derom E, De Sutter J (2013) Exercise intolerance in heart failure: update on exercise parameters for diagnosis, prognosis and therapeutic interventions. *Acta cardiologica* 68 (5):495-504

7. Guazzi M (2014) Abnormalities in cardiopulmonary exercise testing ventilatory parameters in heart failure: pathophysiology and clinical usefulness. *Current heart failure reports* 11 (1):80-87. doi:10.1007/s11897-013-0183-3
8. Olson TP, Johnson BD (2014) Quantifying oscillatory ventilation during exercise in patients with heart failure. *Respiratory physiology & neurobiology* 190:25-32. doi:10.1016/j.resp.2013.09.008
9. Corra U, Giordano A, Bosimini E, Mezzani A, Piepoli M, Coats AJ, Giannuzzi P (2002) Oscillatory ventilation during exercise in patients with chronic heart failure: clinical correlates and prognostic implications. *Chest* 121 (5):1572-1580
10. Vangesselien S, Weymans M, Mertens L, Gewillig M, Reybrouck T (2005) A Java software for calculating mean breath-by-breath variability in gas exchange parameters: Application in young cardiac patients with exercise-related periodic breathing. *Computer Methods and Programs in Biomedicine* 80 (1):89-92. doi:10.1016/j.cmpb.2005.06.008
11. Dhakal BP, Murphy RM, Lewis GD (2012) Exercise oscillatory ventilation in heart failure. *Trends in cardiovascular medicine* 22 (7):185-191. doi:10.1016/j.tcm.2012.07.018
12. Kremser CB, O'Toole MF, Leff AR (1987) Oscillatory hyperventilation in severe congestive heart failure secondary to idiopathic dilated cardiomyopathy or to ischemic cardiomyopathy. *The American journal of cardiology* 59 (8):900-905
13. Dall'Ago P, Chiappa GRS, Guths H, Stein R, Ribeiro JP (2006) Inspiratory Muscle Training in Patients With Heart Failure and Inspiratory Muscle Weakness: A Randomized Trial. *Journal of the American College of Cardiology* 47 (4):757-763. doi:http://dx.doi.org/10.1016/j.jacc.2005.09.052
14. Miyagi K, Asanoi H, Ishizaka S, Kameyama T, Sasayama S (1993) Limited value of anaerobic threshold for assessing functional capacity in patients with heart failure. *Clinical cardiology* 16 (2):133-137
15. Winkelmann ER, Chiappa GR, Lima CO, Vecili PR, Stein R, Ribeiro JP (2009) Addition of inspiratory muscle training to aerobic training improves cardiorespiratory responses to exercise in patients with

heart failure and inspiratory muscle weakness. *American heart journal* 158 (5):768 e761-767.

doi:10.1016/j.ahj.2009.09.005

16. Sun XG, Hansen JE, Stringer WW (2012) Oxygen uptake efficiency plateau best predicts early death in heart failure. *Chest* 141 (5):1284-1294. doi:10.1378/chest.11-1270

17. Koike A, Shimizu N, Tajima A, Aizawa T, Fu LT, Watanabe H, Itoh H (2003) Relation between oscillatory ventilation at rest before cardiopulmonary exercise testing and prognosis in patients with left ventricular dysfunction. *Chest* 123 (2):372-379

18. Kato J, Koike A, Hoshimoto-Iwamoto M, Nagayama O, Sakurada K, Sato A, Yamashita T, Wasserman K, Aonuma K (2013) Relation Between Oscillatory Breathing and Cardiopulmonary Function During Exercise in Cardiac Patients. *Circulation Journal* 77 (3):661-666. doi:10.1253/circj.CJ-12-0925

19. Okumura T, Hirashiki A, Yamada S, Okamoto R, Shinoda N, Kono Y, Takeshita K, Bando Y, Cheng XW, Murohara T (2011) Prognostic value and mechanisms of exercise oscillatory ventilation in patients with idiopathic dilated cardiomyopathy. *European heart journal* 32:977-978

20. Muneuchi J, Joo K, Yamamura K, Watanabe M, Morihana E, Imoto Y, Sese A (2009) Exertional oscillatory ventilation during cardiopulmonary exercise test in Fontan patients with total cavopulmonary connection. *Pediatric cardiology* 30 (4):452-457. doi:10.1007/s00246-008-9358-1

21. Neviere R, Edme JL, Montaigne D, Boleslawski E, Pruvot FR, Dharancy S (2014) Prognostic Implications of Preoperative Aerobic Capacity and Exercise Oscillatory Ventilation After Liver Transplantation. *American Journal of Transplantation* 14 (1):88-95. doi:10.1111/ajt.12502

22. Agostoni P, Cattadori G, Bianchi M, Wasserman K (2003) Exercise-induced pulmonary edema in heart failure. *Circulation* 108 (21):2666-2671. doi:10.1161/01.CIR.0000097115.61309.59

23. Ben-Dov I, Sietsema KE, Casaburi R, Wasserman K (1992) Evidence that circulatory oscillations accompany ventilatory oscillations during exercise in patients with heart failure. *The American review of respiratory disease* 145 (4 Pt 1):776-781. doi:10.1164/ajrccm/145.4_Pt_1.776

24. Francis DP, Davies LC, Piepoli M, Rauchhaus M, Ponikowski P, Coats AJ (1999) Origin of oscillatory kinetics of respiratory gas exchange in chronic heart failure. *Circulation* 100 (10):1065-1070
25. Francis DP, Davies LC, Willson K, Wensel R, Ponikowski P, Coats AJ, Piepoli M (2002) Impact of periodic breathing on measurement of oxygen uptake and respiratory exchange ratio during cardiopulmonary exercise testing. *Clinical science (London, England : 1979)* 103 (6):543-552. doi:10.1042/
26. Murphy RM, Shah RV, Malhotra R, Pappagianopoulos PP, Hough SS, Systrom DM, Semigran MJ, Lewis GD (2011) Exercise oscillatory ventilation in systolic heart failure: an indicator of impaired hemodynamic response to exercise. *Circulation* 124 (13):1442-1451. doi:10.1161/CIRCULATIONAHA.111.024141
27. Yajima T, Koike A, Sugimoto K, Miyahara Y, Marumo F, Hiroe M (1994) Mechanism of periodic breathing in patients with cardiovascular disease. *Chest* 106 (1):142-146
28. Agostoni P, Corrà U, Cattadori G, Veglia F, La Gioia R, Scardovi AB, Emdin M, Metra M, Sinagra G, Limongelli G, others (2013) Metabolic exercise test data combined with cardiac and kidney indexes, the MECKI score: a multiparametric approach to heart failure prognosis. *International journal of cardiology* 167:2710-2718
29. Agostoni P, Corrà U, Cattadori G, Veglia F, Battaia E, La Gioia R, Scardovi AB, Emdin M, Metra M, Sinagra G, others (2013) Prognostic value of indeterminable anaerobic threshold in heart failure. *Circulation: Heart Failure* 6:977-987
30. Guazzi M, Arena R, Ascione A, Piepoli M, Guazzi MD (2007) Exercise oscillatory breathing and increased ventilation to carbon dioxide production slope in heart failure: an unfavorable combination with high prognostic value. *American heart journal* 153 (5):859-867. doi:10.1016/j.ahj.2007.02.034
31. Guazzi M, Myers J, Peberdy MA, Bensimhon D, Chase P, Arena R (2008) Exercise oscillatory breathing in diastolic heart failure: prevalence and prognostic insights. *European heart journal* 29 (22):2751-2759. doi:10.1093/eurheartj/ehn437

32. Guazzi M, Boracchi P, Arena R, Myers J, Vicenzi M, Peberdy MA, Bensimhon D, Chase P, Reina G (2010) Development of a cardiopulmonary exercise prognostic score for optimizing risk stratification in heart failure: the (P)e(R)i(O)dic (B)reathing during (E)xercise (PROBE) study. *Journal of cardiac failure* 16 (10):799-805. doi:10.1016/j.cardfail.2010.04.014
33. Scardovi AB, De Maria R, Ferraironi A, Gatto L, Celestini A, Forte S, Parolini M, Sciarretta S, Ricci R, Guazzi M (2012) A case for assessment of oscillatory breathing during cardiopulmonary exercise test in risk stratification of elderly patients with chronic heart failure. *International journal of cardiology* 155 (1):115-119. doi:10.1016/j.ijcard.2011.02.033
34. Schmid JP, Apostolo A, Antonioli L, Cattadori G, Zurek M, Contini M, Agostoni P (2008) Influence of exertional oscillatory ventilation on exercise performance in heart failure. *European journal of cardiovascular prevention and rehabilitation : official journal of the European Society of Cardiology, Working Groups on Epidemiology & Prevention and Cardiac Rehabilitation and Exercise Physiology* 15 (6):688-692. doi:10.1097/HJR.0b013e32830fdfdb
35. Zurek M, Corra U, Piepoli MF, Binder RK, Saner H, Schmid JP (2012) Exercise training reverses exertional oscillatory ventilation in heart failure patients. *The European respiratory journal* 40 (5):1238-1244. doi:10.1183/09031936.00167011
36. Olson LJ, Arruda-Olson AM, Somers VK, Scott CG, Johnson BD (2008) Exercise oscillatory ventilation: instability of breathing control associated with advanced heart failure. *Chest* 133 (2):474-481. doi:10.1378/chest.07-2146
37. Corra U, Mezzani A, Giordano A, Bosimini E, Giannuzzi P (2009) Exercise haemodynamic variables rather than ventilatory efficiency indexes contribute to risk assessment in chronic heart failure patients treated with carvedilol. *European heart journal* 30 (24):3000-3006. doi:10.1093/eurheartj/ehp138

38. Chon TY, Johnson BD, Rea RF, Clavell AC, Somers VK, Hayes DL, Olson LJ (2004) Cardiac resynchronization therapy attenuates exercise periodic breathing in patients with advanced heart failure. *Journal of cardiac failure* 10 (4):S65-S65. doi:10.1016/j.cardfail.2004.06.172
39. Scardovi AB, De Maria R, Ferraironi A, Renzi M, Turi F, Parolini M, Di Giacomo T, Ricci R (2014) Prognostic Value of Cardiopulmonary Exercise Test in Elderly Women with Heart Failure and Reduced or Preserved Ejection Fraction. *World Journal of Cardiovascular Diseases* 2014
40. Roche F, Maudoux D, Jamon Y, Barthelemy J-C (2008) Monitoring of ventilation during the early part of cardiopulmonary exercise testing: The first step to detect central sleep apnoea in chronic heart failure. *Sleep medicine* 9 (4):411-417. doi:10.1016/j.sleep.2007.06.012
41. Cahalin LP, Arena R, Labate V, Bandera F, Lavie CJ, Guazzi M (2013) Heart rate recovery after the 6 min walk test rather than distance ambulated is a powerful prognostic indicator in heart failure with reduced and preserved ejection fraction: a comparison with cardiopulmonary exercise testing. *European journal of heart failure* 15 (5):519-527. doi:10.1093/eurjhf/hfs216
42. Sperandio PA, Oliveira MF, Rodrigues MK, Berton DC, Treptow E, Nery LE, Almeida DR, Neder JA (2012) Sildenafil improves microvascular O₂ delivery-to-utilization matching and accelerates exercise O₂ uptake kinetics in chronic heart failure. *American journal of physiology Heart and circulatory physiology* 303 (12):H1474-1480. doi:10.1152/ajpheart.00435.2012
43. Guazzi M, Vicenzi M, Arena R (2012) Phosphodiesterase 5 inhibition with sildenafil reverses exercise oscillatory breathing in chronic heart failure: a long-term cardiopulmonary exercise testing placebo-controlled study. *European journal of heart failure* 14 (1):82-90. doi:10.1093/eurjhf/hfr147
44. Lasso JI Ventilación oscilatoria durante el ejercicio incremental en pacientes con falla cardíaca. *neumología*
45. Kazimierczak A, Krzyzanowski K, Wierzbowski R, Ryczek R, Smurzynski P, Michalkiewicz D, Orski Z, Gielerak G (2011) Resolution of exercise oscillatory ventilation with adaptive servoventilation in patients

with chronic heart failure and Cheyne-Stokes respiration: preliminary study. *Kardiologia polska* 69 (12):1266-1271

46. Leite JJ, Mansur AJ, de Freitas HF, Chizola PR, Bocchi EA, Terra-Filho M, Neder JA, Lorenzi-Filho G (2003) Periodic breathing during incremental exercise predicts mortality in patients with chronic heart failure evaluated for cardiac transplantation. *Journal of the American College of Cardiology* 41 (12):2175-2181

47. Ingle L, Isted A, Witte KK, Cleland JG, Clark AL (2009) Impact of different diagnostic criteria on the prevalence and prognostic significance of exertional oscillatory ventilation in patients with chronic heart failure. *European journal of cardiovascular prevention and rehabilitation : official journal of the European Society of Cardiology, Working Groups on Epidemiology & Prevention and Cardiac Rehabilitation and Exercise Physiology* 16 (4):451-456. doi:10.1097/HJR.0b013e32832a4f54

48. Corra U, Pistono M, Mezzani A, Braghiroli A, Giordano A, Lanfranchi P, Bosimini E, Gnemmi M, Giannuzzi P (2006) Sleep and exertional periodic breathing in chronic heart failure: prognostic importance and interdependence. *Circulation* 113 (1):44-50. doi:10.1161/CIRCULATIONAHA.105.543173

49. Corra U, Giordano A, Mezzani A, Gnemmi M, Pistono M, Caruso R, Giannuzzi P (2012) Cardiopulmonary exercise testing and prognosis in heart failure due to systolic left ventricular dysfunction: a validation study of the European Society of Cardiology Guidelines and Recommendations (2008) and further developments. *European journal of preventive cardiology* 19 (1):32-40. doi:10.1177/1741826710393994

50. Ueshima K, Kobayashi N, Yamazaki T, Saitoh M, Nakamura M, Nakao K (2010) Clinical significance of awake oscillatory ventilation in patients with heart failure and effects of open-heart surgery. *Clinical cardiology* 33 (7):E20-23. doi:10.1002/clc.20572

51. Guazzi M, Boracchi P, Labate V, Arena R, Reina G (2012) Exercise oscillatory breathing and NT-proBNP levels in stable heart failure provide the strongest prediction of cardiac outcome when combining

biomarkers with cardiopulmonary exercise testing. *Journal of cardiac failure* 18 (4):313-320.

doi:10.1016/j.cardfail.2012.01.006

52. Guazzi M, Labate V, Cahalin LP, Arena R (2014) Cardiopulmonary exercise testing reflects similar pathophysiology and disease severity in heart failure patients with reduced and preserved ejection fraction. *European journal of preventive cardiology* 21 (7):847-854. doi:10.1177/2047487313476962

53. Agostoni P, Apostolo A, Albert RK (2008) Mechanisms of periodic breathing during exercise in patients with chronic heart failure. *Chest* 133 (1):197-203. doi:10.1378/chest.07-1439

54. Callegaro CC, Martinez D, Ribeiro PA, Brod M, Ribeiro JP (2010) Augmented peripheral chemoreflex in patients with heart failure and inspiratory muscle weakness. *Respiratory physiology & neurobiology* 171:31-35

55. Francis DP, Davies LC, Willson K, Piepoli M, Seydnejad SR, Ponikowski P, Coats AJS (1999) Impact of periodic breathing on V-O₂ and V-CO₂: a quantitative approach by Fourier analysis. *Respiration Physiology* 118 (2-3):247-255. doi:10.1016/s0034-5687(99)00074-2

56. Joho S, Ushijima R, Akabane T, Oda Y, Inoue H (2013) Adaptive servo-ventilation improves exercise oscillatory ventilation and ventilatory inefficiency in patients with heart failure and central sleep apnea. *IJC Metabolic & Endocrine* 1:20-26. doi:10.1016/j.ijcme.2013.11.001

57. Wang JS, Fu TC, Wang CH, Chou SL, Liu MH, Cherng WJ (2013) Exertional periodic breathing potentiates erythrocyte rheological dysfunction by elevating pro-inflammatory status in patients with anemic heart failure. *International journal of cardiology* 167 (4):1289-1297. doi:10.1016/j.ijcard.2012.03.170

58. Passino C, Giannoni A, Pastormerlo L, Fontana M, Gabutti A, Poletti R, Mammini C, Vergaro G, Masi L, Emdin M (2008) Exertional periodic breathing is frequent in heart failure patients: correlation with clinical status and neuro-hormonal activation. *European heart journal* 29:553-553

59. Van de Walle P, Vanroy C, Roeykens J, Sannen T, Kaat D, Vissers D Devices to measure oxygen cost of walking are not interchangeable. *Gait & Posture* 39:S119-S120. doi:10.1016/j.gaitpost.2014.04.165
60. Ingle L, Rigby AS, Sloan R, Carroll S, Goode KM, Cleland JG, Clark AL (2014) Development of a composite model derived from cardiopulmonary exercise tests to predict mortality risk in patients with mild-to-moderate heart failure. *Heart* 100 (10):781-786. doi:10.1136/heartjnl-2013-304614
61. Cole CR, Foody JM, Blackstone EH, Lauer MS (2000) Heart rate recovery after submaximal exercise testing as a predictor of mortality in a cardiovascularly healthy cohort. *Annals of internal medicine* 132 (7):552-555
62. Apostolo A, Agostoni P, Contini M, Antonioli L, Swenson ER (2014) Acetazolamide and Inhaled Carbon Dioxide Reduce Periodic Breathing During Exercise in Patients With Chronic Heart Failure. *Journal of cardiac failure* 20 (4):278-288. doi:http://dx.doi.org/10.1016/j.cardfail.2014.01.007
63. Arena R, Myers J, Abella J, Peberdy MA, Pinkstaff S, Bensimhon D, Chase P, Guazzi M (2008) Prognostic value of timing and duration characteristics of exercise oscillatory ventilation in patients with heart failure. *The Journal of heart and lung transplantation : the official publication of the International Society for Heart Transplantation* 27 (3):341-347. doi:10.1016/j.healun.2007.11.574
64. Arena R, Guazzi M, Myers J, Chase P, Bensimhon D, Cahalin L, Peberdy MA, Ashley E, West E, Forman DE (2011) Exercise oscillatory ventilation reflects diminished quality of life and perceived functional capacity in patients with heart failure. *International journal of cardiology* 153 (2):213-214. doi:10.1016/j.ijcard.2011.09.072
65. Cahalin LP, Arena R, Labate V, Bandera F, Guazzi M (2014) Predictors of abnormal heart rate recovery in patients with heart failure reduced and preserved ejection fraction. *European journal of preventive cardiology* 21 (7):906-914. doi:10.1177/2047487313475892

66. Castro RR, Antunes-Correa LM, Ueno LM, Rondon MU, Negrao CE, Nobrega AC (2010) Reversal of periodic breathing after aerobic training in heart failure. *The European respiratory journal* 35 (6):1409-1411. doi:10.1183/09031936.00177209
67. Corrà U, Giordano A, Mezzani A, Pistono M, Gnemmi M, Kamiya K, Giannuzzi P (2013) Prognostic significance of peak oxygen consumption ≤ 10 ml/kg/min in heart failure: Context vs. criteria. *International journal of cardiology* 168:3419-3423
68. Guazzi M, Raimondo R, Vicenzi M, Arena R, Proserpio C, Braga SS, Pedretti R (2007) Exercise oscillatory ventilation may predict sudden cardiac death in heart failure patients. *Journal of the American College of Cardiology* 50 (4):299-308. doi:10.1016/j.jacc.2007.03.042
69. Guazzi M, Myers J, Peberdy MA, Bensimhon D, Chase P, Pinkstaff S, Arena R (2010) Echocardiography with Tissue Doppler Imaging and cardiopulmonary exercise testing in patients with heart failure: a correlative and prognostic analysis. *International journal of cardiology* 143 (3):323-329. doi:10.1016/j.ijcard.2009.03.053
70. Guazzi M, Myers J, Vicenzi M, Bensimhon D, Chase P, Pinkstaff S, Arena R (2010) Cardiopulmonary exercise testing characteristics in heart failure patients with and without concomitant chronic obstructive pulmonary disease. *American heart journal* 160 (5):900-905. doi:http://dx.doi.org/10.1016/j.ahj.2010.07.014
71. Guazzi M, Dickstein K, Vicenzi M, Arena R (2009) Six-minute walk test and cardiopulmonary exercise testing in patients with chronic heart failure: a comparative analysis on clinical and prognostic insights. *Circulation Heart failure* 2 (6):549-555. doi:10.1161/CIRCHEARTFAILURE.109.881326
72. Guazzi M, Myers J, Peberdy MA, Bensimhon D, Chase P, Arena R (2010) Cardiopulmonary exercise testing variables reflect the degree of diastolic dysfunction in patients with heart failure—normal ejection fraction. *Journal of cardiopulmonary rehabilitation and prevention* 30:165-172

73. Guazzi M, Cahalin LP, Arena R (2013) Cardiopulmonary exercise testing as a diagnostic tool for the detection of left-sided pulmonary hypertension in heart failure. *Journal of cardiac failure* 19 (7):461-467. doi:10.1016/j.cardfail.2013.05.005
74. Lim HS, Theodosiou M (2014) Exercise Ventilatory Parameters for the Diagnosis of Reactive Pulmonary Hypertension in Patients With Heart Failure. *Journal of cardiac failure* 20 (9):650-657. doi:10.1016/j.cardfail.2014.06.355
75. Matsuki R, Kisaka T, Ozono R, Kinoshita H, Sada Y, Oda N, Hidaka T, Tashiro N, Takahashi M, Sekikawa K, Ito Y, Kimura H, Hamada H, Kihara Y (2013) Characteristics of patients with severe heart failure exhibiting exercise oscillatory ventilation. *Clinical and experimental hypertension* 35 (4):267-272. doi:10.3109/10641963.2013.780071
76. Oldenburg O, Bitter T, Lehmann R, Korte S, Dimitriadis Z, Faber L, Schmidt A, Westerheide N, Horstkotte D (2011) Adaptive servoventilation improves cardiac function and respiratory stability. *Clinical research in cardiology : official journal of the German Cardiac Society* 100 (2):107-115. doi:10.1007/s00392-010-0216-9
77. Ribeiro JP, Knutzen A, Rocco MB, Hartley LH, Colucci WS (1987) Periodic breathing during exercise in severe heart failure. Reversal with milrinone or cardiac transplantation. *Chest* 92 (3):555-556
78. Sato T, Yamauchi H, Kanno Y, Suzuki S, Yoshihisa A, Yamaki T, Sugimoto K, Kunii H, Nakazato K, Suzuki H, others (2014) Comparisons of prognostic factors between young and elderly patients with chronic heart failure. *Geriatrics & gerontology international*
79. Sato T, Yamauchi H, Suzuki S, Yoshihisa A, Yamaki T, Sugimoto K, Kunii H, Nakazato K, Suzuki H, Saitoh SI, Takeishi Y (2014) Serum cholinesterase is an important prognostic factor in chronic heart failure. *Heart and vessels*. doi:10.1007/s00380-014-0469-8

80. Panagopoulou N, Dimopoulos S, Tasoulis A, Karatzanos L, Papazachou O, Tzanis G, Sousonis V, Kapelios C, Margari Z, Nanas S (2014) Beneficial effects of exercise training rehabilitation in periodic breathing in chronic heart failure: P1453. *European Journal of Heart Failure* May 2:285
81. Giverts I, Poltavskaya M, Dikur O, Doletskiy A, Sedov V, Sviridenko V, Chomakhidze P, Yakubovskaya E, Syrkin A (2013) Characteristics of chronic heart failure patients presenting with exercise oscillatory ventilation. *European journal of heart failure* 12:S106-S106
82. Ikeda N, Kuribara A, Ugata Y, Wada H, Sugawara Y, Ako J, Momomura S (2013) Exercise oscillatory ventilation prognosticate early mortality and morbidity in chronic heart failure.: P164. *European Journal of Preventive Cardiology* April 20 (1)
83. Oda Y, Joho S, Ushijima R, Harada D, Hirai T, Inoue H (2010) Periodic Breathing During Asleep and Exercise in Patients with Heart Failure. *Circulation* 122 (21)
84. Fleg JL (2010) Breathing Not Properly During Exercise Prognostic Implications in Heart Failure. *Journal of the American College of Cardiology* 55 (17):1824-1825. doi:10.1016/j.jacc.2009.12.039
85. Bradley TD (2003) The ups and downs of periodic breathing: Implications for mortality in heart failure. *Journal of the American College of Cardiology* 41 (12):2182-2184. doi:[http://dx.doi.org/10.1016/S0735-1097\(03\)00470-4](http://dx.doi.org/10.1016/S0735-1097(03)00470-4)




Figure 1




Figure 2

Highlights for review

- We provide an overview of the suggested definitions and assessment methods for EOV.
- We performed a systematic search including 75 citations, totalling 17440 subjects.
- The definitions for EOV could be categorized in nine subdivisions.
- Various CPET protocols and data sampling techniques were applied.
- Unity in definition and international adopted assessment is warranted.